

National Institute for Environmental Studies

Certificate of Analysis

NIES CRM No. 10-d Rice Flour-Unpolished

This environmental certified reference material (CRM), which continues NIES CRM No.10 series, was developed and certified by the National Institute for Environmental Studies (NIES) for the determination of cadmium and other elements in unpolished rice and materials of similar matrix. Certified values are given for eight elements and reference values for a further five elements.

Certified Values

Element	Mass fraction		Analytical method *
	Unit	Certified value Uncertainty	
Magnesium (Mg)	%	0.156 0.007	ICP-MS, ICP-OES
Potassium (K)	%	0.384 0.024	AAS, ICP-MS, ICP-OES
Cadmium (Cd)	mg/kg	0.401 0.034	AAS, ICP-MS, ICP-OES, ID-ICP-MS
Copper (Cu)	mg/kg	7.18 0.63	ICP-MS, ICP-OES
Iron (Fe)	mg/kg	16.4 1.8	ICP-MS, ICP-OES
Manganese (Mn)	mg/kg	9.59 0.82	ICP-MS, ICP-OES
Strontium (Sr)	mg/kg	0.217 0.024	ICP-MS, ICP-OES
Zinc (Zn)	mg/kg	26.5 1.1	ICP-MS, ICP-OES

All certified values were determined based on dry mass. The uncertainty attached to the certified values is the expanded uncertainty using a coverage factor $k = 2$, corresponding to the half-width of a confidence interval of approximately 95 %.

* AAS, atomic absorption spectroscopy

ICP-MS, inductively coupled plasma-mass spectrometry

ICP-OES, inductively coupled plasma-optical emission spectrometry

ID-ICP-MS, isotope dilution-inductively coupled plasma-mass spectrometry

Reference Values

Element	Mass fraction		Analytical method *
	Unit	Reference value	
Phosphorus (P)	%	0.386	AAS, ICP-OES

Sulfur (S)	%	0.128	ICP-OES
Calcium (Ca)	mg/kg	107	ICP-OES
Molybdenum (Mo)	mg/kg	0.762	ICP-MS, ICP-OES
Sodium (Na)	mg/kg	13.5	AAS, ICP-MS, ICP-OES

All reference values were determined based on dry mass.

* AAS, atomic absorption spectroscopy

ICP-MS, inductively coupled plasma-mass spectrometry

ICP-OES, inductively coupled plasma-optical emission spectrometry

Characterization

The property values of the material were statistically determined based on chemical analyses by 8 organizations (including 14 laboratories) using a wide range of methods. A property value satisfying the following conditions was accepted as a certified value:

- 1) the relative standard deviation associated with the mean of the laboratory means was 5 % or less,
- 2) the number of laboratories contributing to the mean of the laboratory means was at least eight, and
- 3) the number of analytical methods contributing to the mean of the laboratory means was at least two.

The uncertainty attached to the certified values is the expanded uncertainty using a coverage factor $k = 2$, corresponding to the half-width of a confidence interval of approximately 95 %. A property value failing to satisfy one or two of the NIES criteria for certification but supplying valuable additional information about the material is given as a reference value. All certified and reference values were determined based on dry mass.

Description of the Material

The CRM is supplied as fine powder in an amber glass bottle. It is off-white in color.

Preparation of the CRM

This CRM was produced from rice that had been hydroponically cultivated under strictly controlled conditions within the Ecosystem Research Field of NIES. The raw material for the CRM was unpolished rice (about 10 kg) that had been harvested by hand. It was dried at 70 °C for 24 h, crushed in a rotor mill with a 200 µm-sieving ring and homogenized with a V-blender. The powdered unpolished rice was placed in amber glass bottles (15 g in each bottle, 609 bottles), and sterilized by ⁶⁰Co irradiation (20 kGy). All procedures complied with ISO Guide 34.

Homogeneity

Mass fractions of multi-elements, including those for which certified values are given, were determined by ICP-OES and ICP-MS in material taken from 10 bottles selected from the total 609 bottles by stratified random sampling. The between-bottle variation evaluated by a one-way analysis of variance (ANOVA) showed the homogeneity standard deviations between bottles for the analytes to be less than 1 %. The material, therefore, is sufficiently homogeneous for its intended use as a reference material.

Stability

Stability tests demonstrated that long-term (6 months) and short-term (2 weeks at -20 °C or 50 °C) variations in the concentrations of all measured elements in the material were insignificant.

Instructions for Use

1. Care should be taken to avoid contamination when opening the bottles. It is desirable to use up the contents as quickly as possible after opening.
2. This CRM should be kept tightly closed in its original bottle and stored in a desiccator at room temperature (≤ 30 °C).
3. Prior to weighing portions for analysis, the contents of the bottle should be shaken gently.
4. It is recommended that a sample intake of 0.2 g is the minimum for convenient handling.
5. Precautions should be taken to avoid inhalation of the material.
6. This CRM should not be used for purposes other than research. When disposing of the material, local laws concerning processing and disposal of waste materials should be strictly adhered to.
7. The mass fractions of elements in this CRM are reported on a dry mass basis. This CRM, as received, has a moisture content of approximately 3 % measured in NIES by drying a separate sub-sample for 4 h at 85 °C. Correction to dry mass should be determined by drying a separate sub-sample at the time of use.

Expiry Date of Certification

The expiry date for the certified values of this CRM is September 2022 assuming that the recommended storage conditions are adhered to. NIES will notify via its website if any changes in the contents are recognized within the term of validity.

Collaborating Laboratories in Analysis

The certified and reference values for this CRM were based on analytical values from the following participating organizations:

National Institute for Environmental Studies; Environmental Control Center Co., Ltd.; IDEA Consultants, Inc.; Japan Food Research Laboratories; Murata Measuring Instrument Service Ltd.; Naitoh Environmental Science Co., Ltd.; Nittech Research Corporation; Shimadzu Techno-Research, Inc.

Technical Information

Technical information and the latest reports regarding this material can be obtained from the website.

<http://www.nies.go.jp/labo/crm-e/index.html>

September 15, 2012

Takashi Imamura

Director

Center for Environmental Measurement and Analysis

National Institute for Environmental Studies

Health and Environmental Risk Division,
National Institute for Environmental Studies,
16-2 Onogawa, Tsukuba, Ibaraki 305-8506 Japan
FAX: +81-29-850-2900, Email: nies.crm@nies.go.jp

Original certificate date: September 15, 2012

Certificate revision date: April 1, 2021 (Editorial changes)

DRAFT